





ADDITIONAL INFORMATION ON THE TENDER CONDITION Nr. 1

Contracting Authority: Fyzikální ústav AV ČR, v.v.i. **Seat:** Na Slovance 2, 182 21 Praha 8

Identification No.: 68378271 **Tax Identification No.:** CZ68378271

Person authorized to represent

the Contracting Authority: Prof. Jan Řídký, DrSc., Director

Public Contract name: MOVPE Apparatus

The contracting entity in accordance with § 49 of Act No. 137/2006 Coll., on Public Procurement, as amended (hereinafter the "Act"), announces the following additional information on the tender conditions relating to the public supply contract published in the Journal of procurement contracts under the registration number 479977.

Question:

Could you tell me please, how often are you planning to execute the MOCVD process and what is the duration of the typical process? It would be enough for me, if you could tell me, how many times a week and how many hours does the typical process take. I know that it would be only an estimation, but I need these data for dimensioning of your column.

Answer:

A typical process will take approx. 6-7 hours. We suppose to use the apparatus 2-3 times a week, depending on requirements.

In Prague	
	prof. Jan Řídký, DrSc., director